CE Declaration of Conformity

Name of Supplier	Status Instruments Ltd
Address:	Status Business Park, Gannaway Lane, Tewkesbury, Gloucestershire.
Country:	United Kingdom

Declare under our sole responsibility that the following apparatus:

Product description:	Exd Flameproof Head
Model or Type No.:	Type 50

Are in conformity with the following relevant EC legislation:

ATEX directive 94/9/EC

Based on the following harmonised standards:

EN 60079-0:2009
EN 60079-1:2007
EN 60079-31:2009

And the following non-harmonised standards:

N/A

And therefore complies with all of the relevant essential requirements of those directives.

The following Notified Body has been involved in the conformity assessment process:

Notified Body	Sira Certification Service
Notified Body No.	0518
Role:	Issue of ATEX EC Type Examination certificate
Certificate No.	Sira 10ATEX1350

Additional information:

ATEX coding

II 2 GD Ex d IIC T6 Gb T_{amb} +55°C Ex t IIIC T85°C Db IP68 T_{amb} +55°C	II 2 GD Ex d IIC T5 Gb T_{amb} +70°C Ex t IIIC T100°C Db IP68 T_{amb} +70°C
II 2 GD Ex d IIC T4 Gb T_{amb} +105°C Ex t IIIC T135°C Db IP68 T_{amb} +105°C	II 2 GD Ex d IIC T3 Gb T_{amb} +170°C Ex t IIIC T200°C Db IP68 T_{amb} +170°C

Name and position of person binding the manufacturer or authorised representative:

Signature

Name	I Pullin
Function	Managing Director
Location	Status Instruments Ltd
Date of issue	01/10/2014