

UK Declaration of Conformity

Product name – SEM1801/2XTC

Description – Atex rail temperature transmitter for thermocouple

Status Instruments Ltd, located at Status Business Park, Gannaway Lane, Tewkesbury, Gloucestershire GL20 8FD, United Kingdom, hereby declares, under the sole responsibility of the manufacturer, that our product complies, by application of the harmonised standards and directives below:

Electromagnetic Compatibility Regulation 2016

BS EN 61326-1:2013 and later amendments

Table 2 industrial environments

RoHS Directive UK Legislation SI 2012 No. 3032

BS EN IEC 63000:2018

ATEX Directive 2014/34/EU & UK SI 2016/1107

Certificate – ExVeritas 25UKEX2064X

Notified Body No. 2585

EN IEC 60079-0: 2018+A11:2024

EN 60079-11:2012

The equipment is marked with the following:

II 1 G D

Ex ia IIC T4 Ga

Ex ia IIIC T 135 DEG C Da

BS EN 60529:1992+A2:2013 Degrees of protection provided by enclosures (IP Code)

Refer to the Product data sheets on the Status Website for specific IP ratings

Signed: 

Title: COMPLIANCE ENGINEER

Signed: 

Title: MANAGING DIRECTOR

(For and on behalf of *Status Instruments Ltd*)

Date: 08.01.2026

Place: Tewkesbury, UK